

for the detection of Antibodies to Treponema pallidum (Tp) in Human Serum / Plasma

INTENDED USE

The TRUSTline Syphilis Ab Rapid Test is a lateral flow chromatographic immunoassay for the qualitative detection of antibodies (IgG, IgM and IgA) to Treponema pallidum (Tp) in human serum or plasma. It is intended to be used by healthcare professionals as a screening test and as an aid in the diagnosis of infection with Tp. The test is not automated and does not require any additional instrument. Any reactive specimen with the TRUSTline Syphilis Ab Rapid Test must be confirmed with alternative testing method(s) and clinical findings

SUMMARY AND EXPLANATION OF THE TEST

Tp, a spirochete bacterium, is the causative agent of the venereal disease syphilis. Although syphilis rates are declining in the United States after an epidemic outbreak between 1986 and 1990¹, the incidence of syphilis in Europe has increased since 1992, especially in the countries of the Russian Federation where peaks of 263 cases per 100,000 have been reported². In 1995, WHO (World Health Organization) reported 12 million new cases of syphilis³. At present, the rate of positive syphilis serological tests among HIV-infected individuals continues to rise.

Serological detection of anti-Tp antibodies has long been recognized as an aid in the diagnosis of syphilis since the natural course of the infection is characterized by periods without clinical manifestations. Both IgM and IgG antibodies were detected in sera from patients with primary and secondary syphilis. The IgM antibody may be detectable towards the second week of an infection while IgG antibodies appear later at approximately 4 weeks⁴. These antibodies can last for several years or even decades in the serum of a patient with untreated latent syphilis5.

Antigens such as Rapid Plasma Reagin (RPR) and Tp bacterial extracts have been used in syphilis serological tests for decades. However, RPR antigen is a non-Treponema antigen derived from bovine heart. Antibodies to RPR antigen do not develop until 1-4 weeks after the appearance of the chancre, thus this antigen lacks sensitivity to primary syphilis. The Tp extracts are prepared from inoculated rabbit testis and contain a certain amount of contaminated materials, such as flagella, which can lead to cross-reactions with borrelia and leptospires in the serological test. In addition, the composition of extracts may vary from lot to lot. Recently, several highly immunogenic Tp specific antigens have been identified and used as an alternative to the traditional antigens with the advantage of having high specificity and reproducibility6-9

The TRUSTline Syphilis Ab Rapid Test was developed to detect antibodies (IgG, IgM and IgA) to recombinant antigens of Tp in serum or plasma within 15 minutes. The test can be performed by minimally trained personnel and without cumbersome laboratory equipment.



TEST PRINCIPLE

The TRUSTline Syphilis Ab Rapid Test is a lateral flow chromatographic immunoassay. The test cassette consists of: 1) a burgundy colored conjugate pad containing recombinant Tp antigens conjugated with colloidal gold (Tp conjugates) and a control antibody conjugated with colloidal gold, and 2) a nitrocellulose membrane strip containing a test line (T line) and a control line (C line). The T line is pre-coated with non-conjugated recombinant Tp antigens, and the C line is pre-coated with a control line antibody.

When an adequate volume of test specimen is dispensed into the sample well of the test cassette, the specimen migrates by capillary action across the cassette. Anti-Tp antibody, if present in the specimen, will bind to the Tp conjugates. The immunocomplex is then captured on the membrane by the pre-coated Tp antigen forming a burgundy colored T line, indicating a Tp antibody positive test result. Absence of the T line suggests a negative result.

The test contains an internal control (C line) which should exhibit a burgundy colored line of the immunocomplex of the control antibodies regardless of color development on the T line. If the C line does not develop, the test result is invalid, and the specimen must be retested with another device.

REAGENTS AND MATERIALS PROVIDED

- Individually sealed foil pouches containing: 1.
 - a. One cassette device
 - b. One desiccant
- 2. Specimen transfer device
- 3. One package insert (instruction for use)

MATERIALS REQUIRED BUT NOT PROVIDED

1. Clock or Timer 2. Disposable gloves

WARNINGS AND PRECAUTIONS

For in Vitro Diagnostic Use

- This package insert must be read completely before performing the test. Failure to 1. follow the insert may lead to inaccurate test results.
- 2. Do not open the sealed pouch unless ready to conduct the assay.
- 3. Do not use the test device if pouch is not intact.
- Do not use expired devices or components. 4.
- Bring all reagents to room temperature (15-30°C) before use. 5.
- 6. Do not use the components of different lots and of any other type of test kit as a substitute for the components in this kit.
- Do not use hemolyzed blood for testing. 7.
- Wear protective clothing and disposable gloves while handling the kit reagents and 8. clinical specimens. Wash hands thoroughly after performing the test.
- 9. Users of this test should follow the US CDC Universal Precautions for prevention of transmission of HIV, HBV and other blood-borne pathogens.
- 10. Do not smoke, drink or eat in areas where specimens or kit reagents are being handled.

- 11. Dispose of all specimens and materials used to perform the test as bio-hazardous waste
- 12. Handle the negative and positive controls in the same manner as the patient specimens
- 13. The test result should be read 15 minutes after a specimen is applied to the sample well of the device. Reading the test result after 20 minutes may give erroneous results.
- 14. Do not perform the test in a room with strong air flow, i.e. an electric fan or strong air conditioning
- 15. Clean up spills thoroughly using appropriate disinfectant.
 - REAGENT PREPARATION AND STORAGE INSTRUCTIONS

All reagents are ready to use as supplied. Store test kit at 1-30°C. If stored at 1-8°C, ensure that all reagents are brought to room temperature before opening. The unopened test device is stable through the expiration date printed on the label, when stored at recommended temperature. Do not freeze the kit or expose the kit to temperatures above 30°C. The test device is sensitive to humidity and heat. Perform the test immediately after removing the test device from the foil pouch.

SPECIMEN COLLECTION AND HANDLING

Consider any materials of human origin as infectious and handle them using standard biosafety procedures.

Plasma

- Collect blood specimen into a lavender, blue or green top collection tube (containing EDTA, citrate or heparin, respectively in Vacutainer®) by Step 1: enipuncture
- Step 2: Separate the plasma by centrifugation.
- Step 3: Carefully withdraw the plasma into a new pre-labeled tube.
- Step 1: Collect blood specimen into a red top collection tube (containing no anticoagulants in Vacutainer®) by venipuncture.
- Step 2: Allow the blood to clot.
- Step 3: Separate the serum by centrifugation.
- Step 4: Carefully withdraw the serum into a new pre-labeled tube.

Test specimens as soon as possible after collecting. Store specimens at 2-8°C if not tested immediately. Specimens can be stored at 2-8°C for up to 5 days. The specimens should be frozen at -20°C for longer storage

Avoid multiple freeze-thaw cycles. Prior to testing, bring frozen specimens to room temperature slowly and mix gently. Specimens containing visible particulate matter should be clarified by centrifugation before testing. Do not use samples demonstrating gross lipemia, gross hemolysis or turbidity in order to avoid interference with result interpretation.

ASSAY PROCEDURE

- Step 1: Bring the specimen and test components to room temperature if refrigerated or frozen. Mix the specimen well prior to assay once thawed.
- Step 2: When ready to test, open the pouch at the notch and remove the device. Place the test device on a clean. flat surface
- Step 3: Be sure to label the device with the specimen ID number.
- Fill the specimen transfer device with the specimen. Holding the specimen Step 4: transfer device vertically, dispense 2 drops (about 60- 90 µL) of specimen into the sample well, making sure there are no air bubbles





Note: Add 1 drop of Saline or Phosphate-Saline buffer (common buffers used in clinics not provided in the kit) to the sample well if flow migration is not observed in the result window within 30 seconds, which could occur with highly viscous specimens.

- Step 5: Set up timer.
- Step 6: Results can be read at 15 minutes. Positive results may be visible in as soon as 1 minute

Do not read result after 20 minutes. To avoid confusion, discard the test device after interpreting the result.

QUALITY CONTROL

- Internal Control: This test contains a built-in control feature, the C line. The C line 1. develops after adding specimen. If the C line does not develop, review the entire procedure and repeat the test with a new device.
- 2. External Control: Good Laboratory Practice recommends using external controls, positive and negative, to ensure the proper performance of the assay, particularly under the following circumstances:
 - A new operator uses the kit, prior to performing testing of specimens. a
 - b. A new lot of test kits is used.
 - A new shipment of test kits is used. C.
 - d. The temperature used during storage of the kit falls outside of 1-30°C.
 - The temperature of the test area falls outside of 15-30°C.
 - To verify a higher than expected frequency of positive or negative results. f.
 - q.
 - To investigate the cause of repeated invalid results.



for the detection of Antibodies to Treponema pallidum (Tp) in Human Serum / Plasma

INTERPRETATION OF ASSAY RESULT

 NEGATIVE RESULT: If only the C line is developed, the test indicates that no detectable anti-*Tp* antibody is present in the specimen. The result is negative or nonreactive.



 POSITIVE RESULT: If both the C and T lines are developed, the test indicates the presence of anti-Tp antibodies in the specimen. The result is positive or reactive.



Specimens with positive results should be confirmed with alternative testing method(s) and clinical findings before a diagnostic decision is made.

 INVALID: If no C line is developed, the assay is invalid regardless of color development on the T line as indicated below. Repeat the assay with a new device.



PERFORMANCE CHARACTERISTICS

1. <u>Clinical Performance</u>

A total of 883 samples from susceptible subjects were tested with the TRUSTline Syphilis Ab Rapid Test in comparison with a commercial Syphilis Ab ELISA test. Comparison for all subjects is shown in the following table.

	TRUSTline Syphi		
ELISA	Positive	Negative	Total
Positive	68	0	68
Negative	2	813	815
Total	70	813	883

Relative Sensitivity:100%, Relative Specificity:99.75%, Overall Agreement:99.77%

2. Cross-Reactivity

The negative specimen was spiked with serum specimens of infectious diseases and then tested according to the standard procedure. The results showed that the **TRUSTline** syphilis Ab Rapid Test had no cross-reaction with the following tested serum specimens of infectious disease.

Specimen	Sample Size	Syphilis Ab Reactivity	
Dengue Positive Serum	10	Negative	
HAV Positive Serum	10	Negative	
HCV Positive Serum	10	Negative	
HIV Positive Serum	10	Negative	
HBsAg Positive Serum	10	Negative	
ANA Positive Serum	5	Negative	
RF Positive Serum(≤2,500 IU/mI)	5	Negative	

3. Precision

Within run and between run precisions have been determined by testing 20 replicates with four category of the specimens: negative, weak, medium and strong positive specimens. The negative, weak, medium and strong positive specimens were correctly identified in all of the tests performed in each run.

4. Interference

Common substances (such as pain and fever medication and blood components) may affect the performance of the TRUSTline Syphilis Ab Rapid Test. This was studied by spiking these substances into four levels of Syphilis Ab standard controls. The results are presented in the following table and demonstrate that at the concentrations tested, the substances studied do not affect the performance of the TURSTline Syphilis Ab Rapid Test.

Note: -: Negative; +: Weak Positive;+++: Strong Positive

Detential Interfering	Syphilis Ab Reactivity			
Substances Spiked	Negative	Weak Positive	Strong Positive	
Control	-	+	+++	
Bilurubin 15 mg/dL	-	+	+++	
Creatinine 5 mg/dL	-	+	+++	
Glucose 120 mg/dL	-	+	+++	
Albumin 5g/L	-	+	+++	
Salicylic Acid 4.34 mmol/L	-	+	+++	
EDTA 3.4 µmol/L	-	+	+++	
Urea 40 mg/dL	-	+	+++	

LIMITATIONS OF TEST

- The Assay Procedure and the Interpretation of Assay Result sections must be followed closely when testing for the presence of anti-*Tp* antibody in serum or plasma from individual subjects. Failure to follow the procedure may lead to inaccurate results.
- The TRUSTline Syphilis Ab Rapid Test is limited to the qualitative detection of anti-Tp antibody in human serum or plasma. The intensity of the test line does not have a linear correlation with the antibody titer in the specimen.
- A negative result for an individual subject indicates the absence of detectable anti-Tp antibodies. However, a negative test result does not preclude the possibility of exposure to or infection with Tp.
- 4. A negative result can occur if the quantity of the anti-Tp antibody present in the specimen is below the detection limits of the assay or if the antibodies that are detected are not present during the stage of disease in which a sample is collected.
- False negative results may arise because of hook effect due to very high titer of antibodies in sample. Repeat the test by using different dilution of same sample.
- Infection may progress rapidly. If the symptoms persist and the result from the TRUSTline Syphilis Ab Rapid Test is non-reactive, it is recommended to test with an alternative method or to re-sample the patient a few weeks later.
- Some specimens containing unusually high titers of heterophile antibodies or rheumatoid factor may affect expected results.
- A positive result indicates a past or present infection. The results obtained with this test should only be interpreted in conjunction with other diagnostic procedures and clinical findings.

REFERENCES

- Centers for Disease Control and Prevention. Chlamydia trachomatis infections: policy guidelines from prevention and control. Morbid. Mortal. Weekly Rep. 1995; 34:53S-74S.
- Tichonova, L., K. Borisenko, H.Ward, A.meheus, et al. Epidemics of syphilis in the Russian Federation: Trends, origins, and priorities for control. Lancet 1997; 350:210-213.
- Gerbase, A. C., J. T. Rowley, D. H. Heymann, S. F. Berkley, and P. Piot. Global prevalence and incidence estimates of selected curable STDs. Sex. Transm. Infect 1998; 74:S12-S16.
- Luger AFH. Serological Diagnosis of Syphilis: Current methods. In: Young H, McMillan A, eds. Immunological diagnosis of sexually transmitted diseases. New York: Marcel Decker, 1988: 249-274.
- Baker-Zander SA, Hook EW 3rd, Bonin P, Handsfield HH, Lukehart SA. Antigens of Treponema pallidum recognized by IgG and IgM antibodies during syphilis in humans. J Infect Dis. 1985; 151(2):264-72.
- Norgard MV, Chamberlain NR, Swancutt MA, Goldberg MS. Cloning and expression of the major 47-kilodalton surface immunogen of treponema pallidum in Escherichia Coli. Infect Immun 1986; 54:500-506.
- Purcell BK, Chamberlain NR, Goldberg MS, Andrews LP, Robinson EJ, Norgard MV, Radolf JD. Molecular cloning and characterization of the15-kilodalton major immunogen of Treponema pallidum. Infect Immun. 1989; 57(12):3708-14
- Bailey MJ, Thomas CM, Cockayne A, Strugnell RA, Penn CW. Cloning and expression of Treponema pallidum antigens in Escherichia coli. J Gen Microbiol 1989; 135 (Pt 9):2365-78.
- Sambri V, Marangoni A, Simone MA, D'Antuono A, Negosanti M, Cevenini R. Evaluation of recomWell Treponema, a novel recombinant antigen-based enzymelinked immunosorbent assay for the diagnosis of syphilis. Clin Microbiol Infect 2001; 7(4):200-5.



Website: www.athenesedx.com